

EC DECLARATION OF CONFORMITY
According to Annex II (exemption of section 4)
of Council Directive 93/42/EEC concerning Medical Devices

Reference No. : 1828C-01
Manufacturer : SIDAPHARM P.C.
Facility Address : 21, Stageiriti & 24, Em. Fili str., GR-543 52, Thessaloniki, Greece
Product : **SIDA-VISC** Sodium Hyaluronate Ophthalmic Solution
 Supralon Sodium Hyaluronate Ophthalmic Solution
 (Annex I)
 GMDN: 35907
Classification : IIb, according to Rule 8, Annex IX of Council Directive 93/42/EEC
Guidelines Applicable : MEDDEV 2.4/1 Rev. 9 - June 2010 Rule 8

This Declaration of Conformity is issued under the sole responsibility of the manufacturer. SIDAPHARM hereby declares that the aforementioned product complies with all the essential requirements of Council Directive 93/42/EEC, concerning Medical Devices, as amended by 2007/47/EC.

The compliance of this applicable quality assurance system has been certified by the "Health Technology Certification", which is a Notified Body, according to Council Directive 93/42/EEC, with identification number 2803.

The present is issued, according to EC Certificate No.: 1828C04210204, whose original expiry date was 26/05/2024, but it has been granted an extension, according to Regulation (EU) 2023/607, till 31/12/2028.

The present replaces any previous declaration has been issued for these products.

For and on behalf of:
SIDAPHARM P.C.

SIDAPHARM P.C.
Medical Disposables
Head Offices: 21, Stageiriti & 24, Em. Fili str.
Thessaloniki GR-54352, Greece
Branch: 6, Laertou str., GR-55535, Greece
Tel: +30 2310 906660 - Fax: +30 2310 989846
VAT Reg. No.: EL997296038
Registration Number: 144520204000

Diana Mochintra
General Manager

Date: 04/04/2024

ANNEX I - Variations

Description	Ref. No
SIDA-VISC 1.0% Sodium Hyaluronate Ophthalmic Solution (1,0 ml)	10000
SIDA-VISC 1.4% Sodium Hyaluronate Ophthalmic Solution (1,0 ml)	10005
SIDA-VISC 1.6% Sodium Hyaluronate Ophthalmic Solution (1,0 ml)	10006
SIDA-VISC 1.8% Sodium Hyaluronate Ophthalmic Solution (1,0 ml)	10002
SIDA-VISC 2.0% Sodium Hyaluronate Ophthalmic Solution (1,0 ml)	10009
SIDA-VISC 3.0% Sodium Hyaluronate Ophthalmic Solution (1,0 ml)	10007
SIDA-VISC 1.0% Sodium Hyaluronate Ophthalmic Solution (1,5 ml)	10010
SIDA-VISC 1.4% Sodium Hyaluronate Ophthalmic Solution (1,5 ml)	10014
SIDA-VISC 1.6% Sodium Hyaluronate Ophthalmic Solution (1,5 ml)	10016
SIDA-VISC 1.8% Sodium Hyaluronate Ophthalmic Solution (1,5 ml)	10018
SIDA-VISC 3.0% Sodium Hyaluronate Ophthalmic Solution (1,5 ml)	10030
Supralon 1.0% Sodium Hyaluronate Ophthalmic Solution	10000
Supralon 3.0% Sodium Hyaluronate Ophthalmic Solution	10007

ANNEX II – Applied Standards

Standard Number	Title
EN ISO 9001:2015	Quality Management Systems-Requirements
EN ISO 13485:2016/AC:2018	Medical devices - Quality management systems - Requirements for regulatory purposes
EN ISO 14971:2019	Medical devices - Application of risk management to medical devices
ISO 24971:2020	Medical devices — Guidance on the application of ISO 14971
EN ISO 11607-1:2020	Packaging for terminally sterilized medical devices -Part 1: Requirements for materials, sterile barrier systems and packaging systems
EN ISO 11607-2:2020	Packaging for terminally sterilized medical devices - Part 2: Validation requirements for forming, sealing and assembly processes
EN 556-1:2001/AC:2006	Sterilization of medical devices – Requirements for medical devices to be designated "STERILE" - Part-1 Requirements for terminally sterilized medical devices
EN ISO 11737-1:2018	Sterilization of medical devices - Microbiological methods - Part 1: Determination of a population of microorganisms on products
EN ISO 11737-2:2020	Sterilization of medical devices - Microbiological methods - Part 2: Tests of sterility performed in the definition, validation and maintenance of a sterilization process
EN 62366-1:2015/A1:2020	Medical devices - Application of usability engineering to medical devices
EN ISO 10993-1:2018	Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process
EN ISO 10993-5:2009	Biological Evaluation of Medical Devices - Part 5, Tests for In vitro Cytotoxicity
EN ISO 10993-10:2021	Biological evaluation of medical devices – Part 10: Tests for irritation and delayed-type hypersensitivity
EN ISO 10993-13:2010	Biological evaluation of medical devices — Part 13: Identification and quantification of degradation products from polymeric medical devices
EN ISO 10993-18:2020	Biological evaluation of medical devices – Part 18: Chemical characterization of materials
EN ISO 15223-1:2021	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General requirements
ISO 20417:2021	Information supplied by the manufacturer of medical devices
EN ISO 15798:2022	Ophthalmic implants – Ophthalmic viscosurgical devices
ISO 11138-2:2017	Sterilization of health care products - Biological indicators - Part 2: Biological indicators for ethylene oxide sterilization processes
ISO 14644-1:2015	Cleanrooms and associated controlled environments — Part 1: Classification of air cleanliness by particle concentration
ISO 14644-2:2015	Cleanrooms and associated controlled environments — Part 2: Monitoring to provide evidence of cleanroom performance related to air cleanliness by particle concentration
ISO 14644-3:2019	Clean rooms and associated controlled environments - Test methods
ISO 14644-4:2022	Clean rooms and associated control environments - Design, construction and start - up
EN ISO 17665-1:2006	Sterilization of health care products- Moist heat - Requirements for the development validation & routine control of sterilization process for medical device
EP/USP	European Pharmacopeia, United states of pharmacopoeia