

## Declaration of Conformity

European medical device regulation (EU) 2017/745

Berneck, 7 November 2024

We hereby declare, on our sole responsibility, that the medical devices

Product	<b>MD1105.08</b> <b>OS4 Surgical platform</b>
Reference number	<b>VC860200, VC860300, and corresponding pedal: VE860010</b>
Classification	<b>IIb, according to annex VIII of the regulation MDR (EU) 2017/745</b>
Device category	<b>MDA 0309 – Phacoemulsification/vitreotomy system</b>
Basic UDI-DI	<b>7630003MD110508UU</b>
Manufacturer	Oertli Instrumente AG, Hafnerwissenstrasse 4, 9442 Berneck, Switzerland
SRN	CH-MF-000016175
Intended purpose:	The full assembly of the device variant "anterior/posterior with endolaser", it includes the following functionalities: <ul style="list-style-type: none"><li>– Irrigation and aspiration (I/A function)</li><li>– Ultrasound phaco (PHACO function)</li><li>– Bipolar diathermy for coagulation in the case of bleeding and coaptation of the conjunctiva during the eye surgery (DIA function)</li><li>– Bipolar diathermic capsulotomy (CAPS function)</li><li>– Bipolar diathermic deep sclerotomy ab interno (HFDS GLAU function)</li><li>– Operation of a vitrectomy instrument (VIT function)</li><li>– Injection and extraction of viscoelastic substances (INJECTION and EXTRACTION functions)</li><li>– Retinal photocoagulation with endolaser (LASER function)</li><li>– Intraocular illumination (LUM function)</li><li>– Maintaining intra-ocular pressure by air (AIR function) and active infusion (GFI function, gas-forced infusion)</li><li>– Fluid/air exchange</li></ul>

meet the applicable provisions, in particular the General Safety and Performance Requirements of the regulation (EU) 2017/745 of the European Parliament and of the Council of 5. April 2017 on medical devices.

The Conformity Assessment Procedure according to MDR (EU) Annex IX, Chapter I, III and Section 4 has been performed and the Notified Body has evaluated the Technical Documentation according to Annex II and Annex III.

EC-Certificate, Certificate registration no.:	<b>244057 MDR2017Q</b>
Notified Body:	<b>DQS Medizinprodukte GmbH, August-Schanz-Strasse 21 60433 Frankfurt am Main, Deutschland</b>
Identification number notified body:	<b>0297</b>
EC-Representative:	<b>Oertli Ophthalmomedic Österreich GmbH, Schwefel 93, 6850 Dornbirn, Austria SRN: AT-AR-000003554</b>
Other relevant Union legislations	<b>RoHS Directive 2011/65/EU</b>

This certificate is valid until 14.12.2027

Oertli Instrumente AG



Norbert Brill  
Person Responsible for Regulatory Compliance  
Head of Research and Development  
Executive Board Member