Form: QMS-S041-W36-F3 (Rev.1.5)



34-14 Maehama, Hiroishi-cho, Gamagori, Aichi 443-0038, Japan
TEL +81-533-67-8895 FAX +81-533-68-1320
URL www.nidek.com
e-mail info@nidek.co.jp

Document No. DOCPM-700EU07

DECLARATION OF CONFORMITY

Manufacturer's name	NIDEK Co., Ltd.	SRN	Not issued yet
Manufacturer's address	34-14 Maehama, Hiroishi-c	cho, Gamago	ri, Aichi 443-0038, Japan
European Representative name	NIDEK S.A.	SRN	FR-AR-000000341
European Representative address	Ecoparc, rue Benjamin Fran	klin, 94370 S	ucy En Brie, FRANCE
Identification of device	PD METER		
Model No.	PM-700		
Classification(Annex VIII, MDR)	Imeasuring function (Rule	13)	
Category (for RoHS)	8		
Basic UDI-DI	4987669109GD		
Common Specification	N/A		
Conformity Assessment procedure	Annex IX		

We herewith declare under our sole responsibility that the above mentioned products meet the provisions of the following EC Council Directive(s)/Regulation. All supporting documentation is retained under the premises of the manufacturer. The supporting technical documentation is also retained at NIDEK S.A., Ecoparc, rue Benjamin Franklin, 94370 Sucy En Brie, FRANCE.

General applicable directive(s)/regulation	Notified Body	Date CE Marking was affixed
REGULATION (EU) 2017/745 OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL of 5 April 2017 on medical devices	DEKRA Certification B.V. Meander 1051, 6825 MJ Arnhem P.O. Box 5185, 6802 ED Arnhem, The Netherlands Certificate No.: 4202074CE01	October 21, 2014 (C) 0344
COUNCIL DIRECTIVE 2011/65/EU concerning restriction of the use of certain hazardous substances.	N/A	October 21, 2014

Place: Aichi, Japan

Effective date: August 30, 2023

Signed by

Katsuaki Tohyama Senior Manager

Person responsible for regulatory compliance

regulatory compliance NIDEK Co., Ltd.

Date of signature: August 28, 2023